



Keysight Technologies Malaysia Sdn Bhd (463532-M)
Bayan Lepas Free Industrial Zone
11900 Penang, Malaysia
Keysight Approved Calibration Provider #71456



5962-0476

Certificate of Calibration

Certificate No: EQ243122-6045767-1

Manufacturer: Keysight Technologies

Model No: U8488A

Options Installed With Specifications: 100

Description: USB Thermocouple Power Sensor

Serial No: MY62330009

Date of Calibration: 05-SEP-2022

Temperature: (23 ± 3) °C

Procedure: AE-P-W10891

Humidity: (40 to 70)% RH

This certifies that the equipment has been calibrated using applicable Keysight Technologies procedures in compliance with a quality management system registered to ISO 9001:2015.

As Received Conditions: Factory tested. No incoming data available.

Action Taken:

- No corrective actions were necessary.

As Shipped Conditions: The measured values of the equipment were observed in specification at the points tested.

Remarks or special requirements:

For inquiries regarding this calibration please contact Keysight Technologies at the above address.


Notes:

1) This calibration report shall not be reproduced, except in full.

Traceability Information: Measurements are traceable to the International System of Units (SI) via national metrology institutes (www.keysight.com/find/NMI) that are signatories to the CIPM Mutual Recognition Arrangement.

Print Date: 06-SEP-2022

Khairul Razmi Bin Zayzan
Quality Manager
Keysight Calibration Provider #71456





Keysight Technologies Malaysia Sdn Bhd (463532-M)
Bayan Lepas Free Industrial Zone
11900 Penang, Malaysia
Keysight Approved Calibration Provider #71456



5962-0476

Certificate of Calibration

Certificate No: EQ243122-6045767-1

Calibration Equipment Used

Date Used: Date equipment used in this calibration

<u>Model Number</u>	<u>Model Description</u>	<u>Equipment ID</u>	<u>Date Used</u>	<u>Cal Due Date</u>
E8361C	Microwave vector network analyzer, 10 MHz to 67 GHz	P0789	05-SEP-2022	08-APR-2023
E5071C	ENA Series Network analyzer	PC7387	05-SEP-2022	11-OCT-2022
N4694A	1.85mm Electronic Cal Kit	P0702	05-SEP-2022	27-DEC-2022
85056A	2.4mm Calibration Kit	P7114	05-SEP-2022	15-APR-2023
E4417A	Pk Pwr Mtr Dual Chn	P0224	05-SEP-2022	07-OCT-2022
N8488A	Average Power Sensor	P0740	05-SEP-2022	03-NOV-2022
E8257D	PSG analog signal generator	PC7398	05-SEP-2022	14-SEP-2022
11667C	Power Splitter	P0736	05-SEP-2022	28-APR-2023
N8488A	Average Power Sensor	P0825	05-SEP-2022	27-JUN-2023
E4419B	Dual Chn Pwr Meter	P0607	05-SEP-2022	23-OCT-2022
N5230A	PNA-L network analyzer	P0370	05-SEP-2022	28-AUG-2023
E4438C	ESG vector signal generator	PA0053	05-SEP-2022	25-OCT-2022
8481A	Power Sensor	P0751	05-SEP-2022	26-OCT-2022
N5222A	PNA Network Analyzer	PA0100	05-SEP-2022	15-SEP-2022
E4417A	Pk Pwr Mtr Dual Chn	PB7358	05-SEP-2022	30-MAR-2023